

#### Phase 1

Class of 2022

1-2020

# Automated Assessment of Cardiothoracic Ratios on Chest Radiographs Using Deep Learning

Varun Danda Thomas Jefferson University, varun.danda@jefferson.edu

Paras Lakhani, MD Thomas Jefferson University, paras.lakhani@jefferson.edu

Follow this and additional works at: https://jdc.jefferson.edu/si\_dh\_2022\_phase1

Part of the Computer Engineering Commons, and the Radiology Commons Let us know how access to this document benefits you

## **Recommended Citation**

Danda, Varun and Lakhani, MD, Paras, "Automated Assessment of Cardiothoracic Ratios on Chest Radiographs Using Deep Learning" (2020). *Phase 1.* Paper 16. https://jdc.jefferson.edu/si\_dh\_2022\_phase1/16

This Article is brought to you for free and open access by the Jefferson Digital Commons. The Jefferson Digital Commons is a service of Thomas Jefferson University's Center for Teaching and Learning (CTL). The Commons is a showcase for Jefferson books and journals, peer-reviewed scholarly publications, unique historical collections from the University archives, and teaching tools. The Jefferson Digital Commons allows researchers and interested readers anywhere in the world to learn about and keep up to date with Jefferson Scholarship. This article has been accepted for inclusion in Phase 1 by an authorized administrator of the Jefferson Digital Commons. For more information, please contact: JeffersonDigitalCommons@jefferson.edu.

# SKMC Class of 2022: SI/DH Abstract

Word count: 300

# Automated Assessment of Cardiothoracic Ratios on Chest Radiographs Using Deep Learning Varun Danda, Paras Lakhani, MD\*

# Introduction:

The cardiothoracic ratio (CTR) is a quantitative measure of cardiac size that can measured from chest radiography (CXR). Although radiologists using digital workstations possess the ability to calculate CTR, clinical demands prevent calculation for every case. In this study, the efficacy of a deep convolutional neural network (dCNN) to assess CTR was evaluated.

### Methods:

611 HIPAA-compliant de-identified CXRs were obtained from [institution blinded] and public databases. Using ImageJ, a board-certified radiologist (reader #1) and a medical student (reader #2), measured the CTR by marking four pixels on all CXRs: the rightand left-most chest wall, the right- and left-most heart border.

The Tensorflow framework (v2.0, Google LLC, Mountain View, CA) and the Keras library (v2.3, https://keras.io) were used to train the dCNN. The images were split into training (511 images), validation (50 images), and test (50 images). U-Net network

1

architecture with an Intersection over Union loss function was employed to predict oval masks on new CXRs and calculate the CTR.

### **Results:**

45 test cases were analyzed. The mean absolute difference in the calculated CTR was 0.026 (stdev: 0.039) for reader 1 vs dCNN, 0.024 (stdev: 0.039) for reader 2 vs. dCNN, and 0.022 (stdev: 0.024) for reader 1 vs. reader 2. The intra-class correlation coefficient was 0.84 (95% CI: 0.73-0.91), 0.84 (95% CI: 0.72-0.91), 0.92 (95% CI: 0.822-0.958) for reader 1 vs. dCNN, reader 2 vs. dCNN, and reader 1 vs. reader 2, respectively.

### **Discussion:**

The dCNN trained in this study outputted similar CTR measurements to the human readers with the dCNN achieving "good" reliability with the human readers and the human readers achieving "excellent" reliability among themselves. This study proves the feasibility of using a dCNN to perform automated CTR assessment from CXR. Future improvements to the algorithm can allow the dCNN to closely approach the expected limits of inter-observer human agreement.

2