

1-2020

Assessment of Dobhoff Tube Malposition on Radiographs Using Deep Learning

Kevin George

Thomas Jefferson University, kevin.george@jefferson.edu

Paras Lakhani, MD

Thomas Jefferson University, paras.lakhani@jefferson.edu

Follow this and additional works at: https://jdc.jefferson.edu/si_dh_2022_phase1



Part of the [Computer Engineering Commons](#), and the [Radiology Commons](#)

[Let us know how access to this document benefits you](#)

Recommended Citation

George, Kevin and Lakhani, MD, Paras, "Assessment of Dobhoff Tube Malposition on Radiographs Using Deep Learning" (2020). *Phase 1*. Paper 15.

https://jdc.jefferson.edu/si_dh_2022_phase1/15

This Article is brought to you for free and open access by the Jefferson Digital Commons. The Jefferson Digital Commons is a service of Thomas Jefferson University's [Center for Teaching and Learning \(CTL\)](#). The Commons is a showcase for Jefferson books and journals, peer-reviewed scholarly publications, unique historical collections from the University archives, and teaching tools. The Jefferson Digital Commons allows researchers and interested readers anywhere in the world to learn about and keep up to date with Jefferson scholarship. This article has been accepted for inclusion in Phase 1 by an authorized administrator of the Jefferson Digital Commons. For more information, please contact: JeffersonDigitalCommons@jefferson.edu.

SKMC Class of 2022: SI/DH Abstract

Word count: 295

Assessment of Dobhoff Tube Malposition on Radiographs Using Deep Learning

Kevin George, Paras C. Lakhani* M.D.

Introduction: Dobhoff tubes (DHT) are narrow-bore flexible devices that deliver enteral nutrition for critically ill patients. Tracheobronchial insertion of DHTs presents a significant risk for pulmonary complications. Thus, DHT insertion requires radiologist confirmation of correct placement with chest x-ray (CXR), increasing clinical delays. To address this, we demonstrate the novel application of Deep Convolutional Neural Networks (DCNNs) to automatically and accurately identify DHTs in CXRs in real time.

Methods: 141 de-identified HIPAA compliant frontal view chest radiographs containing DHTs in various positions were obtained. The DHTs were first manually segmented and verified by a board certified radiologist. Images were split into training (126) and test (15) sets. Data augmentation consisted of horizontal flipping, rotation, sheer, and translation steps. A pretrained deep convolutional neural network model with the U-Net architecture was employed. This net was trained using TensorFlow 2.0 and a 1080ti NVIDIA GPU. The training ran for 300 epochs with an Adam optimizer (learning rate = 0.0001), using an intersection over union (IOU) loss function.

Results: The fully trained network achieved a Sørensen–Dice coefficient of 0.7 between the predicted and ground truth segmentations. This suggests that the DCNN

was able to identify DHT both accurately and in a variety of use cases. Run time per image was less than a second, demonstrating the efficiency of this computer-based method.

Discussion: A Dice coefficient of 0.7 represents strong accuracy and supports the hypothesis that DCNN may be employed to automatically identify DHT positioning. This suggests that deep learning can segment and highlight DHTs, potentially aiding clinical teams. Performance could improve with more training cases and standardization of preprocessing. Future directions include research on the real world impact of such solutions on clinical teams, including whether such a system improves safe DHT placement outcomes on floors.

No references/tables/figures/appendices allowed.