

2010

Motion Control Consistency with Abdominal Compression Device from Daily 4D CBCT

Y. Cui

Thomas Jefferson University and Hospitals

A. S. Harrison

Thomas Jefferson University and Hospitals

Y. Xiao

Thomas Jefferson University and Hospitals

Follow this and additional works at: <https://jdc.jefferson.edu/bodinejournal>



Part of the [Oncology Commons](#)

[Let us know how access to this document benefits you](#)

Recommended Citation

Cui, Y.; Harrison, A. S.; and Xiao, Y. (2010) "Motion Control Consistency with Abdominal Compression Device from Daily 4D CBCT," *Bodine Journal*: Vol. 3: Iss. 1, Article 10.

DOI: <https://doi.org/10.29046/TBJ.003.1.009>

Available at: <https://jdc.jefferson.edu/bodinejournal/vol3/iss1/10>

This Article is brought to you for free and open access by the Jefferson Digital Commons. The Jefferson Digital Commons is a service of Thomas Jefferson University's [Center for Teaching and Learning \(CTL\)](#). The Commons is a showcase for Jefferson books and journals, peer-reviewed scholarly publications, unique historical collections from the University archives, and teaching tools. The Jefferson Digital Commons allows researchers and interested readers anywhere in the world to learn about and keep up to date with Jefferson scholarship. This article has been accepted for inclusion in *Bodine Journal* by an authorized administrator of the Jefferson Digital Commons. For more information, please contact: JeffersonDigitalCommons@jefferson.edu.

Motion Control Consistency with Abdominal Compression Device from Daily 4D CBCT

Cui, Y., Harrison, A.S., Xiao, Y.

Department of Radiation Oncology, Thomas Jefferson University and Hospitals, Philadelphia, PA

Purpose

To investigate the effectiveness of abdominal compression in tumor motion control for lung cancer radiotherapy. To study the motion control consistency during radiation treatment course through 4D reconstruction of daily cone beam computed tomography (CBCT) images.

Method and Materials

Patients with lung tumors were scanned with 4D helical CT and the tumor motion was evaluated in left-right (LR), superior-inferior (SI), and anterior-posterior (AP) directions. Patients with tumor motion greater than 1cm in any of the three directions and who could not tolerate the active breathing control (ABC) were treated with abdominal compression. Kilovoltage(kV) CBCT image was acquired for each treatment fraction to correct patient setup. The CBCT imaging data were then reconstructed using a respiratory-correlated technique to derive 4D CBCT images with 10 different phases of breathing cycle. Daily tumor motion with abdominal compression was evaluated from the 4D CBCT by recording the tumor center position relative to the treatment isocenter in each phase. Three patient data (1 upper lobe tumor and 2 middle lobe tumors) were included in this study and 4 consecutive treatment fractions for each patient were studied. The extent of tumor motion was compared between different treatment fractions to assess the daily variation of motion control with abdominal compression.

Results

Mean value and standard deviation (SD) of tumor motions from all patients and treatment fractions studied were $3.1\text{mm} \pm 0.9\text{mm}$ (mean \pm SD), $6.8\text{mm} \pm 1.7\text{mm}$, and $5.3\text{mm} \pm 1.8\text{mm}$ in LR, SI, and AP directions respectively. The variation of tumor motion extent of the same patient between different fractions was $1.1\text{mm} \pm 0.7\text{mm}$, $2.0\text{mm} \pm 1.9\text{mm}$, and $2.3\text{mm} \pm 2.0\text{mm}$, in LR, SI, and AP directions respectively.

Conclusion

Tumor motions which were greater than 1cm with free-breathing were reduced to less than 1cm with abdominal compression. Variation of motion control during treatment course was significant relative to the tumor motion in the patients and treatment fractions studied.