

10-24-2022


Cannabidiol, Δ^9 -Tetrahydrocannabinol, and Metabolites in Human Blood by Volumetric Absorptive Microsampling and LC-MS/MS Following Controlled Administration in Epilepsy Patients

Federica Pigliasco
University of Genoa

Sara Malaca
Università Politecnica delle Marche

Alfredo Fabrizio Lo Faro
Università Politecnica delle Marche

Anastasio Tini
Follow this and additional works at: <https://jdc.jefferson.edu/iehpfp>
Università Politecnica delle Marche

 Part of the [Internal Medicine Commons](#), and the [Neurology Commons](#)

Giuliana Cangemi

IRCCS Istituto Giannina Gaslini

[Let us know how access to this document benefits you](#)

Recommended Citation

See next page for additional authors

Pigliasco, Federica; Malaca, Sara; Lo Faro, Alfredo Fabrizio; Tini, Anastasio; Cangemi, Giuliana; Cafaro, Alessia; Barco, Sebastiano; Riva, Antonella; Pisati, Angelica; Amadori, Elisabetta; Striano, Pasquale; Tagliabracchi, Adriano; Huestis, Marilyn A.; and Busardò, Francesco Paolo, "Cannabidiol, Δ^9 -Tetrahydrocannabinol, and Metabolites in Human Blood by Volumetric Absorptive Microsampling and LC-MS/MS Following Controlled Administration in Epilepsy Patients" (2022). *Institute of Emerging Health Professions Faculty Papers*. Paper 17.
<https://jdc.jefferson.edu/iehpfp/17>

This Article is brought to you for free and open access by the Jefferson Digital Commons. The Jefferson Digital Commons is a service of Thomas Jefferson University's [Center for Teaching and Learning \(CTL\)](#). The Commons is a showcase for Jefferson books and journals, peer-reviewed scholarly publications, unique historical collections from the University archives, and teaching tools. The Jefferson Digital Commons allows researchers and interested readers anywhere in the world to learn about and keep up to date with Jefferson scholarship. This article has been accepted for inclusion in Institute of Emerging Health Professions Faculty Papers by an authorized administrator of the Jefferson Digital Commons. For more information, please contact: JeffersonDigitalCommons@jefferson.edu.

Authors

Federica Pigliasco, Sara Malaca, Alfredo Fabrizio Lo Faro, Anastasio Tini, Giuliana Cangemi, Alessia Cafaro, Sebastiano Barco, Antonella Riva, Angelica Pisati, Elisabetta Amadori, Pasquale Striano, Adriano Tagliabracci, Marilyn A. Huestis, and Francesco Paolo Busardò



OPEN ACCESS

EDITED BY

Francesca Baratta,
University of Turin, Italy

REVIEWED BY

Abuzar Kabir,
Florida International University,
United States
Carmina Sirignano,
University of Naples Federico II, Italy

*CORRESPONDENCE

Sebastiano Barco,
sebastianobarco@gaslini.org

[†]These authors have contributed equally to this work

SPECIALTY SECTION

This article was submitted to Experimental Pharmacology and Drug Discovery, a section of the journal Frontiers in Pharmacology

RECEIVED 07 September 2022

ACCEPTED 10 October 2022

PUBLISHED 24 October 2022

CITATION

Pigliasco F, Malaca S, Lo Faro AF, Tini A, Cangemi G, Cafaro A, Barco S, Riva A, Pisati A, Amadori E, Striano P, Tagliabracchi A, Huestis MA and Busardò FP (2022), Cannabidiol, Δ^9 -tetrahydrocannabinol, and metabolites in human blood by volumetric absorptive microsampling and LC-MS/MS following controlled administration in epilepsy patients. *Front. Pharmacol.* 13:1038754. doi: 10.3389/fphar.2022.1038754

COPYRIGHT

© 2022 Pigliasco, Malaca, Lo Faro, Tini, Cangemi, Cafaro, Barco, Riva, Pisati, Amadori, Striano, Tagliabracchi, Huestis and Busardò. This is an open-access article distributed under the terms of the Creative Commons Attribution License (CC BY). The use, distribution or reproduction in other forums is permitted, provided the original author(s) and the copyright owner(s) are credited and that the original publication in this journal is cited, in accordance with accepted academic practice. No use, distribution or reproduction is permitted which does not comply with these terms.

Cannabidiol, Δ^9 -tetrahydrocannabinol, and metabolites in human blood by volumetric absorptive microsampling and LC-MS/MS following controlled administration in epilepsy patients

Federica Pigliasco^{1†}, Sara Malaca^{2†}, Alfredo Fabrizio Lo Faro², Anastasio Tini², Giuliana Cangemi³, Alessia Cafaro^{3,4}, Sebastiano Barco^{3*}, Antonella Riva¹, Angelica Pisati¹, Elisabetta Amadori¹, Pasquale Striano^{1,5}, Adriano Tagliabracchi², Marilyn Ann Huestis⁶ and Francesco Paolo Busardò²

¹Department of Neurosciences, Rehabilitation, Ophthalmology, Genetics, Maternal and Child Health (DINO GMI), University of Genoa, Genoa, Italy, ²Department of Excellence-Biomedical Sciences and Public Health, Università Politecnica delle Marche, Ancona, Italy, ³Chromatography and Mass Spectrometry Section, Central Laboratory of Analyses, IRCCS Istituto Giannina Gaslini, Genoa, Italy, ⁴Department of Internal Medicine, Pharmacology & Toxicology Unit, University of Genoa, Genoa, Italy, ⁵Paediatric Neurology and Muscular Disease Unit, IRCCS Istituto Giannina Gaslini, Genoa, Italy, ⁶Institute of Emerging Health Professions, Thomas Jefferson University, Philadelphia, PA, United States

Cannabidiol (CBD) exhibits anti-inflammatory, anxiolytic, antiseizure, and neuroprotective properties without addictive or psychotropic side effects, as opposed to Δ^9 -tetrahydrocannabinol (THC). While recreational cannabis contains higher THC and lower CBD concentrations, medical cannabis contains THC and CBD in different ratios, along with minor phytocannabinoids, terpenes, flavonoids and other chemicals. A volumetric absorptive microsampling (VAMS) method combined with ultra-high-performance liquid chromatography coupled with mass spectrometry in tandem for quantification of CBD, THC and their respective metabolites: cannabidiol-7-oic acid (7-COOH-CBD); 7-hydroxy-cannabidiol (7-OH-CBD); 6- α -hydroxy-cannabidiol (6- α -OH-CBD); and 6- β -hydroxycannabidiol (6- β -OH-CBD); 11-Hydroxy- Δ^9 -tetrahydrocannabinol (11-OH-THC) and 11-Nor-9-carboxy- Δ^9 -tetrahydrocannabinol (THCCOOH). After overnight enzymatic glucuronide hydrolysis at 37°C, samples underwent acidic along with basic liquid-liquid extraction with hexane: ethyl acetate (9:1, v/v). Chromatographic separation was carried out on a C18 column, with the mass spectrometer operated in multiple reaction monitoring mode and negative electrospray ionization. Seven patients with intractable epilepsy were dosed with various CBD-containing formulations and blood collected

just before their daily morning administration. The method was validated following international guidelines in toxicology. Linear ranges were (ng/ml) 0.5–25 THC, 11-OH-THC, THCCOOH, 6- α -OH-CBD and 6- β -OH-CBD; 10–500 CBD and 7-OH-CBD; and 20–5000 7-COOH-CBD. 7-COOH-CBD was present in the highest concentrations, followed by 7-OH-CBD and CBD. This analytical method is useful for investigating CBD, THC and their major metabolites in epilepsy patients treated with CBD preparations employing a minimally invasive microsampling technique requiring only 30 μ L blood.

KEYWORDS

cannabinoids, medical cannabis, serum, CBD metabolites, UHPLC-MS/MS

1 Introduction

The most researched phytocannabinoids are cannabidiol (CBD) and Δ^9 -tetrahydrocannabinol (THC) (Degenhardt et al., 2017; Barco et al., 2018). CBD exhibits anti-inflammatory, antiseizure, anxiolytic, and neuroprotective properties without addictive or psychotropic effects, as opposed to THC (Pigliasco et al., 2020). While recreational cannabis generally contains high THC concentrations, medical cannabis contains THC and CBD in varying amounts, along with minor terpenes, flavonoids, phytocannabinoids, and other chemicals (Malaca et al., 2021).

There is growing interest in cannabis-based therapies and clinical applications (Lattanzi et al., 2019, 2020; Arzimanoglu et al., 2020). Different CBD products are available on the market with effects varying based on purity, formulation, and concentration. The European Medicines Agency has authorized Epidyolex[®], a pure CBD oral solution. As an add-on therapy for drug-resistant epilepsies including Dravet syndrome, tuberous sclerosis complex and Lennox-Gastaut syndrome (Dubois et al., 2020).

Therapeutic drug monitoring (TDM) of a variety of antiepileptic medications (ASMs) is critical in therapeutic management of patients with epilepsy. TDM is particularly useful in emerging clinical practice for cannabis-based therapies to identify the dose associated with an optimal response (Patsalos et al., 2018; Striano et al., 2008; Brandt, 2019). Although TDM is often performed on plasma or serum samples, a major challenge is a need for repeated venipunctures, which is stressful, especially for children. Dried blood spots and other microsampling techniques offer advantages including easy, rapid, and less invasive sample collection, low sample volumes of 10–50 μ L, minimal sample preparation and safe sample handling with minimum risk of transmission of infectious diseases (Biagini et al., 2020).

Other analytical methods focused on CBD and metabolite identification and quantification in different biological matrices, but none included simultaneous quantification with the volumetric absorptive microsampling (VAMS) method (Barco et al., 2018; Dubois et al., 2020; Pérez-Acevedo et al., 2020, 2021; Pichini et al., 2020, 2021; Pigliasco et al., 2020; Busardò et al., 2021; Malaca et al., 2021). This method quantifies CBD, THC, cannabidiol-7-oic acid (7-COOH-CBD); 7-hydroxy-cannabidiol

(7-OH-CBD); 6- α -hydroxy-cannabidiol (6- α -OH-CBD); and 6- β -hydroxycannabidiol (6- β -OH-CBD); 11-Hydroxy- Δ^9 -tetrahydrocannabinol (11-OH-THC) and 11-Nor-9-carboxy- Δ^9 -tetrahydrocannabinol (THCCOOH) in whole blood collected with VAMS and analyzed with ultra-high performance liquid chromatography coupled with mass spectrometry in tandem (UHPLC-MS/MS). In our previous research, we demonstrated that CBD plasma concentrations were comparable to those measured in venous or capillary blood with VAMS, allowing the use of this microsampling fingerpick device (Barco et al., 2017; D'Urso et al., 2019; Pigliasco et al., 2020). The validated method quantified cannabinoids in children's blood with drug-resistant epilepsy.

2 Materials and methods

2.1 Chemicals and reagents

Standards for CBD, 7-OH-CBD, 7-COOH-CBD, 6- α -OH-CBD and 6- β -OH-CBD, THC, 11-OH-THC, and THCCOOH were obtained from Dalton Research Molecules (Toronto, ON, Canada) and deuterated internal standards (ISTD) THC- d_3 , 11-OH-THC- d_3 , THCCOOH- d_3 and CBD- d_3 were purchased from Cayman Chemical (Ann Arbor, MI, United States) and stored at -20°C until use. LC-MS grade water, LC grade acetone, formic acid and acetonitrile acquired from Sigma-Aldrich[®] (Milano, Italy). Ammonium formate 5 mM was prepared with 97% pure ammonium formate ammonium salt (Sigma-Aldrich[®]) dissolved in LC-MS grade water. Beta-glucuronidase from *E. Coli* (>20,000 units mg/protein) was obtained from Sigma Aldrich[®] (Milano, Italy).

2.2 Instrumental conditions for UHPLC-MS/MS

A Waters[®] Xevo[®] TQ-S micro mass spectrometer (triple quadrupole), equipped with an electrospray ionization source operating in both negative and positive-ion mode (ESI), was used

TABLE 1 Mass spectrometry parameters for analytes and internal standards.

Compounds	Internal Standard	Cone voltage (eV)	Q1 mass (m/z)	Quantification transition		Confirmation transition		RT (min)
				Q3 mass (m/z)	CE (eV)	Q3 mass (m/z)	CE (eV)	
Standards								
7-COOH-CBD	11-OH-THC-d ₃	40	343.1	179.2	20	231.2	26	5.57
6- α -OH-CBD	11-OH-THC-d ₃	30	329.2	158.2	32	173.1	28	6.17
7-OH-CBD	11-OH-THC-d ₃	25	329.1	261.2	20	268.1	24	6.42
6- β -OH-CBD	11-OH-THC-d ₃	30	329.2	158.2	26	173.2	34	6.50
THCCOOH	THCCOOH-d ₃	40	345.1	193.1	24	299.2	24	7.61
11-OH-THC	11-OH-THC-d ₃	30	331.2	193.1	24	201.1	24	8.23
CBD	CBD-d ₃	45	315.2	123	34	189.1	22	9.79
THC	THC-d ₃	45	315.2	123	34	193.1	22	11.27
Internal Standards								
THCCOOH-d ₃	-	40	348.1	196.2	26	-	-	7.61
11-OH-THC-d ₃	-	30	334.2	196.1	30	-	-	8.22
CBD-d ₃	-	45	316.1	110.1	45	248.2	45	9.77
THC-d ₃	-	45	318.2	123	34	196.1	22	11.27

CBD, Cannabidiol; 7-COOH-CBD, 7-Carboxy-cannabidiol; 7-OH-CBD, 7-Hydroxy-cannabidiol; 6- α -OH-CBD, 6- α -Hydroxy-cannabidiol; 6- β -OH-CBD, 6- β -Hydroxy-cannabidiol; THC, Δ^9 -Tetrahydrocannabinol; 11-OH-THC, 11-Hydroxy- Δ^9 -tetrahydrocannabinol; THCCOOH, 11-Nor-9-carboxy- Δ^9 -tetrahydrocannabinol; CE, collision energy; RE, retention time.

to conduct the UHPLC-MS/MS analysis. The instrument was interfaced with an ACQUITY UPLC[®] I-Class (Waters[®]; Milan, Italy). Data were collected using the MassLynx[®] program version 4.1 (Waters[®], Milano, Italy). A Waters[®] ACQUITY[™] PREMIER UPLC[®] BEH C18 column (100 × 2.1 mm, 1.7 μ m) was used for separation. Run time was 17 min with mobile phases (A) water with ammonium formate 5 mM pH 7.5 and (B) acetonitrile at a flow rate of 0.4 ml/min. The gradient program went from 5% B for 0.25 min to 30% B after 1 min, 80% B after 11.5 min and held for 0.5 min, 100% B after 11.55 min till 13.5 min, and then back to 5% B after 13.55 min and held for the remaining 17:00 min. Column oven and autosampler temperatures were 50°C and 10°C, respectively. The mass spectrometer was operated in multiple reaction monitoring (MRM) mode, with two transitions for each analyte and ISTD (see Table 1). By individually injecting neat standards into methanol and ramping cone voltage and collision energy, MS parameter settings were made to be fully optimized. (see Table 1). The ESI conditions were optimized to source temperature 150°C, capillary voltage -2.8 kV, cone gas flow rate 0.18 ml/min, desolvation temperature 650°C, and desolvation gas flow rate 1200 L/h. The scan speed (dwell time) was 0.023 s.

2.3 Preparation of quality control samples and calibration standards

Standard stock solutions with all five non-deuterated standards were prepared in methanol at 1 mg/ml, 100 μ g/

ml, 10 μ g/ml, and 1 μ g/ml. ISTD stock solution with THC-d₃, THCCOOH-d₃, 11-OH-THC-d₃ and CBD-d₃ was prepared in methanol at 1 μ g/ml. Due to the unavailability of deuterated standards for the analytes, the deuterated 11-OH-THC standard was used for the CBD metabolites. Glass vials containing the stock solutions were kept at -20°C.

Pre-screened blood samples were donated by Department of Neurosciences, Rehabilitation, Ophthalmology, Genetics, Maternal and Child Health (DI-NOGMI, University of Genoa, Genoa, Italy) as material discarded during current clinical practice. During method validation, samples were evaluated to rule out any potential sources of chromatographic interferences and then mixed to form a homogeneous pool of blank samples for the preparation of calibration standards and quality control samples.

Calibrator working solutions were prepared by diluting standard stock solutions with methanol (0.25, 0.5, 1, 2, 5 and 25 ng/ml for THC, 11-OH-THC, THCCOOH, 6- α -OH-CBD and 6- β -OH-CBD; 1.5, 25, 50, 100, 250 and 500 ng/ml for CBD and 7-OH-CBD; 3.5, 20, 100, 500, 1000 and 5000 ng/ml for 7-COOH-CBD). Low, medium and high-quality control (QC) working solutions were prepared daily from the standard stock solutions in methanol. THC, 11-OH-THC, THCCOOH, 6- α -OH-CBD and 6- β -OH-CBD QC concentrations were 0.75, 1.5 and 20 ng/ml, CBD and 7-OH-CBD QC concentrations were 5, 80 and 400 ng/ml and 7-COOH-CBD QC concentrations were 10, 350 and 4000 ng/ml.

2.4 Sample preparation

Since prior research and our preliminary experiments in real samples revealed that CBD metabolites were present as glucuronides, blood samples were extracted following enzymatic hydrolysis (Bergamaschi et al., 2013; Ujváry and Hanuš, 2016). Glucuronide hydrolysis was carried out by adding 2 μ L ISTD solution (100 ng/ml), beta-glucuronidase (*E. Coli* >20,000 units mg/protein) was added to 30 ml of blood in a VAMS tip, followed by dilution in 800 ml of water and heated overnight at 37°C. After hydrolysis, acidic and basic extractions were performed to extract all metabolites based on their acid-base properties. For the basic extraction, 100 μ L ammonium hydroxide (pH 9) was added to the hydrolyzed sample after the addition of 4 ml hexane: ethyl acetate (9:1). Samples underwent horizontal agitation for 30 min. Samples were centrifuged for 3 min at 3500 rpm, and the supernatant was then transferred to a clean tube. The samples were centrifuged for 3 minutes at 3500 rpm and the supernatant was then transferred to a clean tube. The remaining aqueous phase was then treated with 15 μ L of formic acid ($\geq 99.9\%$) (pH 3) for acid extraction. Four mL of hexane: ethyl acetate (9: 1) was added to the sample. The tubes were then capped, vortexed for 10 s, mixed for 10 min and centrifuged at 5000 g for 5 min. Both extractions' supernatants were combined in a glass tube (Safe-Lock Tube[®], Eppendorf, Milano, Italy) and dried under nitrogen. Samples were resuspended with 100 μ L water: methanol (1:1), transferred into autosampler glass vials, and 10 μ L injected onto the chromatographic system.

2.5 Validation of the analytical method

2.5.1 Sensitivity

Sensitivity was determined by analyzing four replicates of negative samples and examination of the signal/noise ratio. The standard deviation (SD) of the mean noise level over the retention time window of analytes was used to determine the detection limit (LOD = 3 SD) and the quantification limit (LLOQ = 10 SD). The calculated LLOQ had to demonstrate precision and accuracy within the 20% relative SD and relative error, respectively, in order to be acceptable.

2.5.2 Selectivity and carryover

Selectivity evaluated the capacity of analytes to be identified in the presence of matrix elements. Blank blood samples were assayed to identify possible endogenous interferences. In addition, blank blood samples were assayed for possible exogenous drug interferences. For this, other commonly encountered analytes (such as common drugs or metabolites) encountered in routine work were analyzed with fortified matrix samples at high therapeutic or lethal concentrations. The acceptance criteria were no signal/noise ratio higher than 3 at ± 0.2 min of the retention time of the analytes (ranging

from 5.57 to 11.27 min) in the quantitative and qualitative ions. To test for carryover, blank blood samples were analyzed immediately after the highest calibrator. Carryover was the highest fortified concentration at which no analyte carryover above the method's LOD was observed in the blank matrix.

2.5.3 Calibration curve

Six calibrators were assayed on five separate days to establish the calibration curve. The peak area ratio of each compound and its corresponding ISTD were plotted against each analyte's concentration. The minimally acceptable linearity requires a coefficient of correlation (r^2) ≥ 0.99 and each calibrator quantifying within $\pm 20\%$ of target concentration. Dilution integrity was checked for over-the-curve samples with concentrations 10 and 50 times higher than the highest calibrators, verifying precision and accuracy to be within 15%.

2.5.4 Imprecision and bias

Imprecision was expressed as the RSD (%), and bias was calculated as (determined concentration)/(nominal concentration) $\times 100\%$. Acceptance criteria for intra- and inter-assay imprecision were CV $\leq 20\%$ and bias $\leq 15\%$. To evaluate intra-assay imprecision, six blank blood samples each were fortified with the target analytes at three different concentrations (low, medium, and high QC) and analyzed on the same day. Evaluation of inter-assay imprecision and bias were performed over 5 days with a minimum of six concentrations.

2.5.5 Matrix effect and recovery

Matrix effect was determined by comparing peak areas of the extracted blank samples fortified with standards after the extraction procedure with the peak areas of pure diluted substances. Recovery was determined by comparing peak area of the extracted compounds fortified before extraction to the peak area obtained from samples fortified post-extraction (representing 100% recovery). The ISTD mixture was added to samples after extraction.

2.5.6 Stability

Compounds' stability in blood was evaluated through repeated analysis ($n = 5$) of QC samples after three freeze-thaw cycles (storage at -20°C) on the compounds stability in blood was evaluated by repeated analysis ($n = 5$) of QC samples. In addition, short term (24 and 48 h) and mid-term (1 month) stability were assessed using five different aliquots of QC stored at -20°C . The stability was expressed as a percentage of the initial concentration (first analyzed batch) of the analytes in QC.

2.6 Application to patients' samples

The analytical method's applicability was demonstrated using real clinical samples from patients taking different CBD

TABLE 2 Patients' demographic, clinical, and treatment data.

Patient ID	Age (y), gender	Weight (Kg)	Epilepsy disorder	CBD formulation, dose (mg/Kg/day), treatment duration (d)	Concomitant drugs
1	9, ♂	24	Dravet Syndrome	Epidyolex oral solution, 20 820	STP, VPA
2	8, ♀	23.6	Aicardi Syndrome	Galenic CBD oil, 3.9, 1125	LTG
3	15, ♀	29	Noonan Syndrome	Galenic CBD oil 24%, 7.2, 240	VPA, LCM, LZP
4	12, ♂	28.4	Focal non-lesional epilepsy	CBD crystals & Bedrolite, 26.4 and 10.6, 2130	LEV, CLB
5	20, ♀	40	Focal non-lesional epilepsy	CBD crystals, 7.5, 1030	VPA, FBM, NTZ
6	3, ♂	11	Infantile spasms/West Syndrome	Epidyolex oral solution, 16.3, 30	VPA, PB, CLB
7	8, ♂	35	Focal non-lesional epilepsy	Galenic CBD oil, 10.8, 850	PB

CLB, clobazam; FBM, felbamate; LEV, levetiracetam; LCM, lacosamide; LTG, lamotrigine; LZP, lorazepam; NTZ, nitrazepam; PB, phenobarbital; STP, stiripentol; VPA, valproate.

formulations. (Epidyolex[®], CBD oil, CBD oil Enecta and CBD crystal) for the treatment of drug-resistant epilepsy of different etiologies (see Table 2). Blood samples were collected in the morning before daily dose administration using the 30 μ L VAMS devices (MITRA[®], Neoteryx, 105 Torrance, CA, United States) for capillary blood collection. Capillary VAMS were obtained in accordance with the manufacturer's recommendations: before pricking the patient's finger with a micro-needle, the area was disinfected and after the first drop of blood was removed, the VAMS tip was placed in contact with the surface of the second drop to absorb the matrix. The study was approved by the Regional Ethical Committee (CER Liguria: 056/057/058/059-2019) and written informed consent was signed by patients or caregivers.

3 Results

Previous analytical methods determined CBD, THC and metabolites (Figure 1) by UHPLC-MS/MS methods with VAMS collection (Dubois et al., 2020; Pigliasco et al., 2020) but no assay is currently available to simultaneously quantify all of these analytes in whole blood.


3.1 Validation of the analytical method

The method was validated over 5 days in blood samples following the most recent criteria for bioanalytical method

development and validation (Peters et al., 2018; Wille et al., 2018), Linearity, sensitivity [limits of detection (LOD) and quantification (LOQ)], selectivity, accuracy, imprecision and carryover were calculated using five daily replicates of calibrators (six for each calibration curve) and five replicates of the three QC samples. Method validation results, presented in Tables 3, 4 were following the internationally established criteria (Peters et al., 2018; Wille et al., 2018). No relevant degradation was observed after any of the three freeze/thaw cycles, with differences in the initial concentration less than 15% for all the compounds under investigation. Similar results (differences from the initial concentration always lower than 15%) were obtained for the case of short-term and mid-term stability tests, confirming the validity of stored samples for analysis.

3.2 Analysis of patients' samples

For proof of concept, the VAMS collection and analytical method were applied to seven samples from seven patients receiving various CBD formulations each at different therapeutic dosages. Four males (ages 3–12 years; weight: 11–28.4 kg) and three females (age range: 8–20 years; weight: 23.6–40 kg) were treated at the Giannina Gaslini Children's Hospital provided samples. The study was approved by the Regional Ethical Committee (CER Liguria: 056/057/058/059-2019) and written informed consent was provided by patients or caregivers. Table 5 summarizes the patients' results.



4 Discussion

An analytical method was validated for the determination of CBD, THC, and their respective metabolites and later applied to clinical samples. Blood samples were collected from seven patients under treatment with CBD formulations. Blood CBD concentrations were higher for the patients treated with Galenic CBD oil (patients #2 and #7)

compared to patients treated with other CBD preparations. 7-COOH-CBD, the inactive metabolite, was present in the highest concentrations, followed by 7-OH-CBD, and CBD. 6- α -OH-CBD and 6- β -OH-CBD concentrations were always lower than concentrations of the other CBD metabolites, but for the first time were detected in all patients' samples, with the highest concentrations in patient #2. In previous studies, these two analytes were undetectable in

TABLE 3 Linearity, limits of detection (LOD) and limits of quantification (LOQ) for analytes under investigation in blood samples.

Analytes	Determination coefficient r^2	LOD ng/mL	LOQ ng/mL
CBD	0.996 ± 0.003	0.50	1.50
7-OH-CBD	0.999 ± 0.002	0.50	1.50
7-COOH-CBD	0.999 ± 0.002	1.10	3.50
6- α -OH-CBD	0.996 ± 0.004	0.10	0.25
6- β -OH-CBD	0.994 ± 0.003	0.10	0.25
THC	0.997 ± 0.002	0.10	0.25
11-OH-THC	0.998 ± 0.002	0.10	0.25
THCCOOH	0.998 ± 0.001	0.10	0.25

CBD, Cannabidiol; 7-COOH-CBD, 7-Carboxy-cannabidiol; 7-OH-CBD, 7-Hydroxy-cannabidiol; 6- α -OH-CBD, 6- α -Hydroxy-cannabidiol; 6- β -OH-CBD, 6- β -Hydroxy-cannabidiol; THC, Δ^9 -Tetrahydrocannabinol; 11-OH-THC, 11-Hydroxy- Δ^9 -tetrahydrocannabinol; THCCOOH, 11-Nor-9-carboxy- Δ^9 -tetrahydrocannabinol; LOD, limit of detection; LOQ, limit of quantification.

TABLE 4 Validation parameters for cannabinoid analytes under investigation in blood samples.

Analytes	Intra-assay accuracy %CV			Inter-assay accuracy %CV			Intra-assay imprecision %CV			Inter-assay imprecision %CV			Recovery (%)		
	L	M	H	L	M	H	L	M	H	L	M	H	L	M	H
7-COOH-CBD	5.1	2.4	5.7	7.9	2.5	4.1	6.9	2.4	4.7	7.0	2.6	4.8	95.7	95.9	97.9
6- α -OH-CBD	7.9	9.1	2.8	7.8	8.4	3.3	3.1	3.5	2.9	7.9	3.0	3.7	96.5	84.3	78.2
7-OH-CBD	7.8	7.5	5.2	6.9	7.7	4.7	8.8	4.5	3.6	6.8	3.1	2.9	93.7	92.2	98.1
6- β -OH-CBD	7.9	5.8	7.8	9.6	8.6	5.2	3.9	2.8	1.5	9.4	6.2	6.5	81.7	87.9	89.9
THC-COOH	4.2	8.3	5.0	5.8	8.6	4.8	5.8	3.9	2.2	5.8	5.8	6.5	83.5	94.0	93.2
11-OH-THC	9.9	10.4	9.1	9.5	8.5	5.5	2.9	3.7	2.7	4.2	3.8	7.0	77.8	86.4	86.4
CBD	10.2	4.1	2.3	7.8	5.8	3.8	9.2	4.8	1.1	6.8	6.4	5.1	69.0	63.2	64.5
THC	7.1	4.9	7.0	6.9	7.4	6.9	5.8	4.0	2.6	6.1	6.7	2.8	80.1	85.6	88.9

CBD, Cannabidiol; 7-COOH-CBD, 7-Carboxy-cannabidiol; 7-OH-CBD, 7-Hydroxy-cannabidiol; 6- α -OH-CBD, 6- α -Hydroxy-cannabidiol; 6- β -OH-CBD, 6- β -Hydroxy-cannabidiol; THC, Δ^9 -Tetrahydrocannabinol; 11-OH-THC, 11-Hydroxy- Δ^9 -tetrahydrocannabinol; THCCOOH, 11-Nor-9-carboxy- Δ^9 -tetrahydrocannabinol; CV, coefficient of variation; L, low; M, medium; H, high.

TABLE 5 Formulations, doses, and concentrations of CBD, THC, and their respective metabolites in patients' blood samples collected with the VAMS technique after CBD therapy.

ID	Formulation	CBD mg/kg/day	Concentration ng/mL							
			6- α -OH-CBD	6- β -OH-CBD	7-OH-CBD	CBD-COOH	CBD	11-OH-THC	THC-COOH	THC
1	Epidyolex oral solution	20	16.8	17.4	399	7335	109	N.D.	N.D.	N.D.
2	Galenic CBD Oil	3.9	28.9	38.8	883	6677	2503	N.D.	N.D.	N.D.
3	CBD Oil 24%	3.9	0.9	1.0	53.5	357	19.1	N.D.	N.D.	N.D.
4	CBD crystal & Bedrolite	250	8.2	5.8	498	15,371	333	N.D.	N.D.	N.D.
5	CBD crystal	100	2.0	4.0	77.3	211	39.0	N.D.	N.D.	N.D.
6	Epidyolex oral solution	0.9	2.6	6.2	266	1120	88.7	N.D.	N.D.	N.D.
7	Galenic CBD oil	18 drops	1.9	1.9	59.4	423	1878	N.D.	N.D.	N.D.

CBD, Cannabidiol; 7-COOH-CBD, 7-Carboxy-cannabidiol; 7-OH-CBD, 7-Hydroxycannabidiol; 6- α -OH-CBD, 6- α -Hydroxycannabidiol; 6- β -OH-CBD, 6- β -Hydroxycannabidiol; 11-OH-THC, 11-Hydroxy- Δ^9 -tetrahydrocannabinol; THCCOOH, 11-Nor-9-carboxy- Δ^9 -tetrahydrocannabinol; THC, Δ^9 -Tetrahydrocannabinol; N.D., not detected.

some patients' samples [4]. THC and its metabolites, THCCOOH and 11-OH-THC, were not detected, indicating that the CBD formulations contained little if any THC.

In conclusion, this simple and rapid UHPLC-MS/MS method enabled robust and sensitive quantification of CBD, THC, and their respective metabolites, with good precision, accuracy, and efficiency. Also, blood sample collection by a microsampling technique (VAMS) is a major advantage when dealing with patients, especially children to avoid invasive procedures during repeated venipunctures.

Data availability statement

The original contributions presented in the study are included in the article/supplementary material, further inquiries can be directed to the corresponding author.

Ethics statement

The studies involving human participants were reviewed and approved by Regional Ethical Committee (CER Liguria: 056/057/058/059-2019). Written informed consent to participate in this study was provided by the participants' legal guardian/next of kin.

Author contributions

Conceptualization, FB, GC, and PS; Validation, SM, AL, AT, and FP; Formal Analysis, AC; Investigation, SM, AL, and AT; Resources, AR, AP, and EA; Writing – Original Draft

Preparation, SM; Writing – Review & Editing, SM, AL, SB, FB, GC, and PS; Supervision, MH, FB, and PS; Funding Acquisition, GC, ADT, and PS.

Funding

This study was funded by the Italian Ministry of Health, RC 2022. We gratefully thank all Italian citizens who allocated the 5 × 1000 share of their tax payment in support of health research.

Acknowledgments

The authors would like to thank Simonetta di Carlo, Antonella Bacosi, Michele Sciotti, Josuè Gottardi, Chiara Fraioli and Laura Martucci for technical assistance.

Conflict of interest

The authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

Publisher's note

All claims expressed in this article are solely those of the authors and do not necessarily represent those of their affiliated organizations, or those of the publisher, the editors and the reviewers. Any product that may be evaluated in this article, or claim that may be made by its manufacturer, is not guaranteed or endorsed by the publisher.

References

- Arzimanoglou, A., Brandl, U., Cross, J. H., Gil-Nagel, A., Lagae, L., Landmark, C. J., et al. (2020). Epilepsy and cannabidiol: A guide to treatment. *Epileptic Disord.* 22, 1–14. doi:10.1684/epd.2020.1141
- Barco, S., Castagnola, E., Moscatelli, A., Rudge, J., Tripodi, G., and Cangemi, G. (2017). Volumetric adsorptive microsampling-liquid chromatography tandem mass spectrometry assay for the simultaneous quantification of four antibiotics in human blood: Method development validation and comparison with dried blood spot. *J. Pharm. Biomed. Anal.* 145, 704–710. doi:10.1016/j.jpba.2017.07.033
- Barco, S., Fucile, C., Manfredini, L., Grandis, E. D., Gherzi, M., Martelli, A., et al. (2018). A UHPLC-MS/MS method for the quantification of Δ^9 -tetrahydrocannabinol and cannabidiol in decoctions and in plasma samples for therapeutic monitoring of medical cannabis. *Bioanalysis* 10, 2003–2014. doi:10.4155/bio-2018-0184
- Bergamaschi, M. M., Barnes, A., Queiroz, R. H. C., Hurd, Y. L., and Huestis, M. A. (2013). Impact of enzymatic and alkaline hydrolysis on CBD concentration in urine. *Anal. Bioanal. Chem.* 405, 4679–4689. doi:10.1007/s00216-013-6837-x
- Biagini, D., Antoni, S., Lomonaco, T., Ghimenti, S., Salvo, P., Bellagambi, F. G., et al. (2020). Micro-extraction by packed sorbent combined with UHPLC-ESI-MS/MS for the determination of prostanoids and isoprostanoids in dried blood spots. *Talanta* 206, 120236. doi:10.1016/j.talanta.2019.120236
- Brandt, C. (2019). Pharmacodynamic monitoring of antiepileptic drug therapy. *Ther. Drug Monit.* 41, 168–173. doi:10.1097/FTD.0000000000000623
- Busardò, F. P., Pérez-Acevedo, A. P., Pacifici, R., Mannocchi, G., Gottardi, M., Papaseit, E., et al. (2021). Disposition of phytocannabinoids, their acidic precursors and their metabolites in biological matrices of healthy individuals treated with vaporized medical cannabis. *Pharmaceuticals* 14, 59. doi:10.3390/ph14010059
- Degenhardt, F., Stehle, F., and Kayser, O. (2017). "Handbook of cannabis and related pathologies biology, pharmacology, diagnosis, and treatment," in *The biosynthesis of cannabinoids*, 13–23. doi:10.1016/B978-0-12-800756-3.00002-8
- Dubois, S., Marchese, F., Pigliasco, F., Barco, S., Tripodi, G., Lomonaco, T., et al. (2020). A volumetric adsorptive microsampling technique to monitor cannabidiol levels in epilepsy patients. *Front. Pharmacol.* 11, 582286. doi:10.3389/fphar.2020.582286
- D'Urso, A., Rudge, J., Patsalos, P. N., and De Grazia, U. (2019). Volumetric adsorptive microsampling: A new sampling tool for therapeutic drug monitoring of antiepileptic drugs. *Ther. Drug Monit.* 41, 681–692. doi:10.1097/FTD.0000000000000652
- Lattanzi, S., Brigo, F., Trinka, E., Zaccara, G., Striano, P., Del Giovane, C., et al. (2020). Adjunctive cannabidiol in patients with Dravet syndrome: A systematic Review and meta-analysis of efficacy and safety. *CNS Drugs* 343 34, 229–241. doi:10.1007/S40263-020-00708-6
- Lattanzi, S., Trinka, E., Russo, E., Striano, P., Citraro, R., Silvestrini, M., et al. (2019). Cannabidiol as adjunctive treatment of seizures associated with Lennox-

Gastaut syndrome and Dravet syndrome. *Drugs Today* 55, 177–196. doi:10.1358/ DOT.2019.55.3.2909248

Malaca, S., Gottardi, M., Pigliasco, F., Barco, S., Cafaro, A., Amadori, E., et al. (2021). UHPLC-MS/MS analysis of cannabidiol and its metabolites in serum of patients with resistant epilepsy treated with CBD formulations. *Pharmaceuticals* 14, 630. doi:10.3390/ph14070630

Patsalos, P. N., Spencer, E. P., and Berry, D. J. (2018). Therapeutic drug monitoring of antiepileptic drugs in epilepsy: A 2018 update. *Ther. Drug Monit.* 40, 526–548. doi:10.1097/FTD.0000000000000546

Pérez-Acevedo, A. P., Busardò, F. P., Pacifici, R., Mannocchi, G., Gottardi, M., Poyatos, L., et al. (2020). Disposition of cannabidiol metabolites in serum and urine from healthy individuals treated with pharmaceutical preparations of medical cannabis. *Pharmaceuticals* 13, E459. doi:10.3390/ph13120459

Pérez-Acevedo, A. P., Pacifici, R., Mannocchi, G., Gottardi, M., Poyatos, L., Papaseit, E., et al. (2021). Disposition of cannabinoids and their metabolites in serum, oral fluid, sweat patch and urine from healthy individuals treated with pharmaceutical preparations of medical cannabis. *Phytother. Res.* 35, 1646–1657. doi:10.1002/ptr.6931

Peters, F. T., Wissenbach, D. K., Busardo, F. P., Marchei, E., and Pichini, S. (2018). Method development in forensic toxicology. *Curr. Pharm. Des.* 23, 5455–5467. doi:10.2174/1381612823666170622113331

Pichini, S., Malaca, S., Gottardi, M., Pérez-Acevedo, A. P., Papaseit, E., Perez-Maña, C., et al. (2021). UHPLC-MS/MS analysis of cannabidiol

metabolites in serum and urine samples. Application to an individual treated with medical cannabis. *Talanta* 223, 121772. doi:10.1016/j.talanta.2020.121772

Pichini, S., Mannocchi, G., Gottardi, M., Pérez-Acevedo, A. P., Poyatos, L., Papaseit, E., et al. (2020). Fast and sensitive UHPLC-MS/MS analysis of cannabinoids and their acid precursors in pharmaceutical preparations of medical cannabis and their metabolites in conventional and non-conventional biological matrices of treated individual. *Talanta* 209, 120537. doi:10.1016/j.talanta.2019.120537

Pigliasco, F., Barco, S., Dubois, S., Marchese, F., Striano, P., Lomonaco, T., et al. (2020). Cannabidiol determination on peripheral capillary blood using a microsampling method and ultra-high-performance liquid chromatography tandem mass spectrometry with on-line sample preparation. *Molecules* 25, E3608. doi:10.3390/molecules25163608

Striano, S., Striano, P., Capone, D., and Pisani, F. (2008). Limited place for plasma monitoring of new antiepileptic drugs in clinical practice. *Med. Sci. Monit.* 14, RA173–RA178.

Ujváry, I., and Hanuš, L. (2016). Human metabolites of cannabidiol: A Review on their formation, biological activity, and relevance in therapy. *Cannabis Cannabinoid Res.* 1, 90–101. doi:10.1089/can.2015.0012

Wille, S. M. R., Coucke, W., De Baere, T., and Peters, F. T. (2018). Update of standard practices for new method validation in forensic toxicology. *Curr. Pharm. Des.* 23, 5442–5454. doi:10.2174/1381612823666170714154444