

Thomas Jefferson University Jefferson Digital Commons

Phase 1 Class of 2022

1-2020

Patient-specific 3D Printed Liver Models for Pre-operative Planning and Improved Patient Adherence

Nathan Ott

Thomas Jefferson University, nathan.ott@jefferson.edu

Miranda Sill

Thomas Jefferson University, miranda.sill@jefferson.edu

Robert S. Pugliese, PharmD, BCPS

Thomas Jefferson University, Robert.Pugliese@jefferson.edu

Ashesh P. Shah, MD

Thomas Jefferson University, ashesh.shah@jefferson.edu

Follow this and additional works at: https://jdc.jefferson.edu/si_des_2022_phase1

Part of the Art and Design Commons, and the Surgery Commons

Let us know how access to this document benefits you

Recommended Citation

Ott, Nathan; Sill, Miranda; Pugliese, PharmD, BCPS, Robert S.; and Shah, MD, Ashesh P., "Patient-specific 3D Printed Liver Models for Pre-operative Planning and Improved Patient Adherence" (2020). *Phase 1*. Paper 10.

https://jdc.jefferson.edu/si_des_2022_phase1/10

This Article is brought to you for free and open access by the Jefferson Digital Commons. The Jefferson Digital Commons is a service of Thomas Jefferson University's Center for Teaching and Learning (CTL). The Commons is a showcase for Jefferson books and journals, peer-reviewed scholarly publications, unique historical collections from the University archives, and teaching tools. The Jefferson Digital Commons allows researchers and interested readers anywhere in the world to learn about and keep up to date with Jefferson scholarship. This article has been accepted for inclusion in Phase 1 by an authorized administrator of the Jefferson Digital Commons. For more information, please contact: JeffersonDigitalCommons@jefferson.edu.

Title: Patient-specific 3D Printed Liver Models for Pre-operative Planning and Improved Patient Adherence **Authors:** Nathan Ott, BS**; Miranda Sill, BS**; Robert S. Pugliese, PharmD, BCPS*; Ashesh Shah, MD

Project Background: 3D anatomical relationships in the liver are not always visually accessible for surgeons performing resections even with advanced imaging options. Firm understanding of these relationships is essential for timely procedures, which can improve patient outcomes and lower hospital expenses. Patient-specific 3D modeling has existed for some time, though it is costly. New cost-effective techniques have surfaced which may yield opportunities for more effective preoperative planning in liver surgery and improved patient adherence.

Methods: Digital patient-specific 3D reconstruction of a liver was completed by interpolating data from MRI scans using 3D Slicer, a segmenting program. The liver model was processed and 3D printed as a shell to be used as a mold. The liver shell, associated vasculature, and tumor were printed using polylactic acid (PLA) filament on an Ultimaker S5 3D printer. Transparent silicone was used as a cast, giving the model a solid form yet still allowing examination of the inside contents.

Results: One completed liver model was used in pre-surgical consultation of a patient with hepatocellular carcinoma undergoing liver resection and during the surgical procedure as a guide for the surgical team. A follow-up survey concerning qualitative aspects of the model administered to the surgical team suggested high accuracy of the model compared to the anatomy observed during the procedure.

Conclusion: Cost-effective techniques in producing patient specific 3D anatomical models appears not only feasible, but highly effective in improving communication between the surgical team during the procedure and also between the surgeon and the patient during pre-surgical consultation. Future research may be conducted concerning the model's visual clarity as well as impact on patient adherence post-op.