

2010

Uncertainty and Margin Study for IMRT, VMAT, and Proton Beam Therapy for Treatment After Radical Prostatectomy

Y. Cui

Thomas Jefferson University Hospitals

A. S. Harrison

Thomas Jefferson University Hospitals

M T. Studenski

Thomas Jefferson University Hospitals

T. N. Showalter

Thomas Jefferson University Hospitals

J. O. Deasy

Washington University, St. Louis

Follow this and additional works at: <https://jdc.jefferson.edu/bodinejournal>



Part of the [Oncology Commons](#)

[See next page for additional authors](#)

[Let us know how access to this document benefits you](#)

Recommended Citation

Cui, Y.; Harrison, A. S.; Studenski, M T.; Showalter, T. N.; Deasy, J. O.; Yu, Y.; Galvin, J. M.; and Xiao, Y. (2010) "Uncertainty and Margin Study for IMRT, VMAT, and Proton Beam Therapy for Treatment After Radical Prostatectomy," *Bodine Journal*: Vol. 3 : Iss. 1 , Article 26.

DOI: <https://doi.org/10.29046/TBJ.003.1.025>

Available at: <https://jdc.jefferson.edu/bodinejournal/vol3/iss1/26>

This Article is brought to you for free and open access by the Jefferson Digital Commons. The Jefferson Digital Commons is a service of Thomas Jefferson University's [Center for Teaching and Learning \(CTL\)](#). The Commons is a showcase for Jefferson books and journals, peer-reviewed scholarly publications, unique historical collections from the University archives, and teaching tools. The Jefferson Digital Commons allows researchers and interested readers anywhere in the world to learn about and keep up to date with Jefferson scholarship. This article has been accepted for inclusion in Bodine Journal by an authorized administrator of the Jefferson Digital Commons. For more information, please contact: JeffersonDigitalCommons@jefferson.edu.

Uncertainty and Margin Study for IMRT, VMAT, and Proton Beam Therapy for Treatment After Radical Prostatectomy

Authors

Y. Cui, A. S. Harrison, M T. Studenski, T. N. Showalter, J. O. Deasy, Y. Yu, J. M. Galvin, and Y. Xiao

Uncertainty and Margin Study for IMRT, VMAT, and Proton Beam Therapy for Treatment After Radical Prostatectomy

Cui, Y.,¹ Harrison, A.S.,¹ Studenski, M.T.,¹ Showalter, T.N.,¹ Deasy, J.O.,² Yu, Y.,¹ Galvin, J.M.,¹ Xiao, Y.¹

¹Department of Radiation Oncology, Thomas Jefferson University Hospitals, Philadelphia, PA

²Washington University, Saint Louis, MO

Purpose/Objective(s)

To compare the uncertainties of 3D dose distributions, caused by the geometrical uncertainty of patient setup, in IMRT, VMAT, and proton plans for post-prostatectomy treatment. To test the effectiveness of a common margin recipe in these three types of treatment plans.

Material/Methods

Four prostate fossa patient datasets were included. For each case, three different plans were carried out: an IMRT plan of nine fields (XiO, Elekta), a VMAT plan, and a proton plan with two lateral active scanning beams (Oncontra, Nucletron). The plan robustness analysis function in CERR (Washington University, St. Louis, MO) software was used to simulate the DVH uncertainty with given systematic (Σ) and random (σ) shifts in three dimensions. Five different combinations of Σ (2-4mm) and σ (2-4mm) representing clinical situations were used for all plans. The DVH uncertainty range (upper and lower bounds) was generated by CERR for each setting of Σ and σ with a certain confidence level (95% was used in this study). We tested CTV coverage using a common margin recipe ($2.5 \Sigma + 0.7 \sigma$) for all IMRT, VMAT, and proton plans.

Results

More than 98% of PTV was covered by 95% of prescription dose in all plans. The upper bound of PTV $V_{95\%}$ was close to 100% in all plans for all Σ and σ settings. The mean values of lower bound of PTV $V_{95\%}$ were 85.4%, 85.0%, and 87.5% for IMRT, VMAT, and proton plans, respectively ($p=0.03$ for IMRT vs. proton, paired samples t -test; $p=0.01$ for VMAT vs. proton; $p=0.36$ for IMRT vs. VMAT). The mean values of ranges (upper minus lower bound) for rectum V_{45Gy} were 7.5% (IMRT), 7.5% (VMAT), 15.6% (proton), and the mean values of ranges for bladder V_{40Gy} were 6.2% (IMRT), 9.2% (VMAT), 12.7% (proton). The proton plans exhibited significantly wider range of rectum and bladder DVHs than the other two treatment techniques ($p<0.05$ for both). Even though the proton plans had lower rectum and bladder doses as compared with IMRT and VMAT, with the uncertainty, the upper bounds were approaching similar doses from IMRT and VMAT. Analysis of PTV $V_{100\%}$, rectum V_{56Gy} , bladder V_{56Gy} showed similar comparison results. The lower bound of CTV

$V_{95\%}$ was larger than 99.4% in all plans with the estimated Σ and σ from the margin recipe, showing the effectiveness of the margin recipe for IMRT, VMAT, and proton plans included in this study.

Conclusion

In this simulation of potential setup uncertainties, larger variation in DVH for bladder and rectum were observed with proton plans than with IMRT and VMAT plans, to the extent that might compromise the advantage of proton plans. The common margin recipe was validated as a method to assure adequate target volume coverage for IMRT, VMAT and proton plans studied.