

1-2020

Ahmed vs Baerveldt Glaucoma Drainage Device in Uveitic Glaucoma

Zoe McWatters

Thomas Jefferson University, zoe.mcwatters@jefferson.edu

Sapna Sinha, MD

Wills Eye Hospital

Allen Ganjei

Drexel University


Marlene Moster, MD

Wills Eye Hospital

Michael Pro, MD

Follow this and additional works at: https://jdc.jefferson.edu/si_ctr_2022_phase1

Wills Eye Hospital

 Part of the [Ophthalmology Commons](#), and the [Translational Medical Research Commons](#)

[Let us know how access to this document benefits you](#)

See next page for additional authors

Recommended Citation

McWatters, Zoe; Sinha, MD, Sapna; Ganjei, Allen; Moster, MD, Marlene; Pro, MD, Michael; Dale, MD, Elizabeth; Lee, MD, Daniel; Myers, MD, Jonathan S.; Katz, MD, L. J.; Fudenberg, MD, Scott; Mantravadi, MD, Anand; Kolomeyer, MD, Natasha; and Razeghnijad, MD, Reza, "Ahmed vs Baerveldt Glaucoma Drainage Device in Uveitic Glaucoma" (2020). *Phase 1*. Paper 34.
https://jdc.jefferson.edu/si_ctr_2022_phase1/34

This Article is brought to you for free and open access by the Jefferson Digital Commons. The Jefferson Digital Commons is a service of Thomas Jefferson University's [Center for Teaching and Learning \(CTL\)](#). The Commons is a showcase for Jefferson books and journals, peer-reviewed scholarly publications, unique historical collections from the University archives, and teaching tools. The Jefferson Digital Commons allows researchers and interested readers anywhere in the world to learn about and keep up to date with Jefferson scholarship. This article has been accepted for inclusion in Phase 1 by an authorized administrator of the Jefferson Digital Commons. For more information, please contact: JeffersonDigitalCommons@jefferson.edu.

Authors

Zoe McWatters; Sapna Sinha, MD; Allen Ganjei; Marlene Moster, MD; Michael Pro, MD; Elizabeth Dale, MD; Daniel Lee, MD; Jonathan S. Myers, MD; L. J. Katz, MD; Scott Fudenberg, MD; Anand Mantravadi, MD; Natasha Kolomeyer, MD; and Reza Razeghnejad, MD

Zoe McWatters

SI/CTR Abstract

Word count: 249 words

***Primary project advisor**

Ahmed vs Baerveldt Glaucoma Drainage Device in Uveitic Glaucoma

Zoe McWatters³; Sapna Sinha, MD¹; Allen Ganjei²; Marlene Moster, MD¹; Michael Pro,
MD¹; Elizabeth Dale, MD¹; Daniel Lee, MD¹; Jonathan S. Myers, MD¹; L. J. Katz, MD¹;
Scott Fudenberg, MD¹; Anand Mantravadi, MD¹; Natasha Kolomeyer, MD¹; Reza
Razeghinejad, MD^{1*}

¹Wills Eye Hospital, Glaucoma Research Center, Philadelphia, PA

²Drexel University College of Medicine, Philadelphia, PA

³Sidney Kimmel Medical College, Thomas Jefferson University, Philadelphia, PA

Introduction: Uveitis is an inflammatory eye disorder which may elevate intraocular pressure (IOP), causing sight-threatening glaucoma. Treatment of refractory uveitic glaucoma involves implantation of a glaucoma drainage device (GDD). Uveitic glaucoma has been a minority diagnosis in prior studies comparing the valved Ahmed GDD and non-valved Baerveldt GDD. Here we compare the safety and efficacy of the Ahmed and Baerveldt GDD in uveitic glaucoma.

Methods: This retrospective comparative study was conducted on patients with uveitic glaucoma (≥ 14 years old) who underwent Ahmed or Baerveldt GDD implantation between 2006–2018 with a minimum follow-up of 3 months. Success was defined as IOP of 6-21 mmHg with (complete success) or without (qualified success) medications, with at least light perception vision and no further glaucoma surgery. Pearson Chi-

squared test, independent t test, and Kaplan-Meier survival model were utilized for statistical analysis.

Results: 137 eyes of 122 patients (67 Ahmed and 70 Baerveldt) were included.

Baseline characteristics were comparable in both groups. The Baerveldt group experienced greater IOP reduction (60.3% vs. 44.5%), higher complete success rate (28.6% vs. 10.4%), higher complication rate (44.3% vs. 20.9%), and higher hypotony rate (10% vs. 0%) than the Ahmed group (all p values <0.05).

Discussion: Our results are in line with pooled analysis of previous GDD comparison studies on various types of glaucoma. For patients with uveitic glaucoma, the Baerveldt implant provided a significantly higher success rate and greater IOP reduction with fewer medications, but was associated with a higher complication rate, compared to the Ahmed implant.